## Progress on Developing a Segmented Detector using ZnS:Ag/<sup>6</sup>Li

### David Reyna Sandia National Laboratories, CA

Sandia National Laboratories is a multi-program laboratory managed and operated by Sandia Corporation, a wholly owned subsidiary of Lockheed Martin Corporation, for the U.S. Department of Energy's National Nuclear Security Administration under contract DE-AC04-94AL85000. SAND2015-10449 C



### **Previous Prototype Deployment**



- As part of a joint SNL/LLNL project to develop aboveground detector technology, we deployed a 4-cell prototype segmented scintillator system
  - Aboveground shielded and unshielded runs in 2011
  - Belowground deployment in 2012-2013
- Very encouraged by performance of Segmented Scintillator prototype
  - This technology is focused on reducing the overall footprint and enabling a transportable detector that can be deployed in highbackground or unshielded locations
  - Demonstrated rejection of backgrounds of 5 orders of magnitude even without an external shield





### **Segmented Scintillator Detector**

- Individual Segments contain organic scintillator with ZnS:Ag/<sup>6</sup>LiF screens on outer surface
  - Tested cells with both plastic and liquid scintillator (plastic preferred)
- Use of ZnS:Ag with <sup>6</sup>LiF allows identification of neutron capture
  - ZnS:Ag is sensitive to alpha from ncapture on Li
  - Very slow scintillator time constant (~100ns) allows pulse shape discrimination to separate n-capture from γ events
- This 4-cell prototype was intended for first testing background rejection only







# **Particle Identification (PID)**

#### **Neutron identification through Pulse Shape Discrimination (PSD)**



#### **Positron Identification through Topology**

- Positrons are rare in nature
  - Deposit most of their kinetic energy very quickly through standard ionization losses
- Positrons will annihilate into two back-to-back 511 keV gammas
  - Very distinctive signature
  - Gammas will travel ~2-5" through most scintillators





### **Aboveground Data**



**David Reyna** 

**Sandia National Laboratories** 

### **WLS Allows Extended Length**

Test Bar	Attenuation length	Normalized neutron efficiency §
Original grease coupled 60cm	35.6 +- 1.2 cm	10.1 +- 0.9 %
WLS, air gap, 60 cm	118 +- 7 cm	10.8 +- 0.9 %
WLS, air gap, 120 cm	154 +- 11 cm	12.6 +- 1.2 %
WLS, air gap, 180 cm	200 +- 20 cm	10.9 +- 1.1 %

# § Efficiency calculated relative to a calibrated <sup>3</sup>He detector and normalized to detector area

More details can be found in *Sweany et al, NIM A Volume 769, 1 January 2015, 37–43z* 





# This Study – Investigate Performance vs. Size Scaling

- Simulated detector configurations
  - Configurations of 3x3 to 11x11 (9 to 121 segments)
  - WLS segments of 180cm length
    - Position and energy smearing applied in post-processing
    - Low energy (100 keV) threshold applied for trigger
- Simulation of fast neutron backgrounds
  - Geant 4.10.1.p01 with QGSP\_BERT\_HP physics list
    - Older 4.9.5.p02 showed unphysical transitions in neutron spectra that made ~30% reduction in background events
  - Cosmic neutron generator from Gordon et al. (2004)
    - Distributed azimuthally like muons
- Simulation of antineutrino events
  - Coincident positron and 200 keV neutron
  - Uniformly distributed throughout





### **Gut-Check: Neutrons**

**Skin Depth:** antinu 10<sup>2</sup> cosmic 10 • arb. units ٦Ů 10-2 120 0 20 40 60 80 100 number of segments

**Neutron Captures in outer most segments** 

- antinu: neutrons started at center
- cosmic: pencil beam pointed at center

Rate normalized to total number of events



**Neutron Wander:** 

Neutron Captures outside of e+ neighbors

- antinu: uniformly distributed
- cosmic: cosmogenic distribution

Rate normalized to total number of events

#### Suggests a neutron / e+ co-location cut will improve with size





### **Gut-Check: Positrons**



Number of events for which the annihilation gammas are not in the 8 surrounding segments from the identified core e+ deposition

- antinu: uniformly distributed
- cosmic: cosmogenic distribution

Rate normalized to total number of events

#### Suggests a full e+ topology cut will improve with size

# Impact of Antineutrino Event Selection

#### **Overall efficiency for uniformly distributed antineutrino events passing an event definition**



For large detectors, the most restrictive cut yields ~10% efficiency



### **Realistic Expectations**



#### **Expected Rates 25m from SONGS**

#### **Sanity Check**

Approximate duplication of 50m deployment at SONGS

- 4-segments (60 cm)
- Reduced efficiency (non-WLS)
- No proximity selection

Cosmic: 7 events/day

Antinu: 1 events/day

Compares favorably with measured 6 events/day

#### Signal scales like mass but background <u>rejection</u> improves quicker



- This technology appears to be robust and scalable
  - Simulations agree with current measurements and confirm expectation that background rejection improves with size
- A 20' shipping container could be a functional system
  - Contain 520 2m-long segments
  - Total mass ~20 tons
  - Overall efficiency ~10% → ~ 2 tons "ideal"
    - Almost reach Huber's "ideal 5t detector"
  - Expected rates would be reasonable
    - Background rate ~400 events/day
    - Signal rate ~5 events/day/MW<sub>th</sub> at 20m



Now we just need to build one and validate

